

Phthalate Analysis

1. Analytical Condition

YL 6500 Series

Oven : 120°C(2min)-> 20°C/min->250°C(15min)

Column : DB-1701 (30m*0.25mm*0.25um)

Carrier gas : N₂, 1ml/min (Split ratio 15:1)

Injector : Capillary 280°C

Detector : FID 280°C

Injection Volume : 2ul

2. Chromatogram

