



Analysis of Raw Materials for Hair Dyes (14 mix)

1. Analytical Condition

YL9100 HPLC

Mobile Phase : Gradient program

(A: 20mM Na₂HPO₄+10mM CH₃COONH₄, B: ACN)

Flow rate : 1.2 mL/min

Column : Capcellpak MG C18 (4.6 * 250 mm, 5 μm)

Detector : UV/Vis, 254 nm

Column Temp. : 35 °C

Injection volumn: 10 μL

<Gradient Program>

Time(min)	A(%)	B(%)
Initial	100	0
5	100	0
20	0	100
35	0	100
35.1	100	0
45	100	0

2. Chromatogram

