

Sulfur Compounds Analysis

1. Analytical Condition

YL 6500 Series

Oven : 70°C (1 min)-20°C/min-200°C(5 min)

Column : GS-Q(30m*0.53mm)

Carrier gas : He, 20ml/min

Injector : TD adsorption (-20°C, 10min) desorption (250°C)

Detector : FPD 240°C

2. Chromatogram

