

Residual Solvents in Pharmaceutical Ingredients

1. Analytical Condition

Residual Solvent Analyzer

Oven : 50°C (1 min) -> 20 °C/min -> 150 °C (0.1min)

Column : HP-INNOWAX (30 m * 0.53 mm * 1 µm)

Carrier gas : He, 8mL/min (Split ratio 10:1)

Injector : Capillary 230 °C

Detector : FID 250 °C

Injection Volume: 0.5 µL

2. Chromatogram

