

# Vitamin K1 Analysis

## 1. Analytical Condition

### YL 9100 HPLC

Mobile Phase : Methanol

Flow Rate : 1.0ml/min

Column Oven : 40°C

Detector : UV/Vis 280nm

Column : C18 (4.6\*250mm, 5um)

Injection Volumn : 20ul sample loop

## 2. Chromatogram

