

General Sugars Analysis

1. Analytical Condition

YL 9100 Series HPLC

Mobile Phase : ACN:Water=75:25(Isocratic mode)

Flow Rate : 1ml/min

Column Oven : 40°C

Detector : ELSD (SC 43 °C /DT 60 °C, Cal 50, Filter=9)

Column : Shodex NH2 (4.6*250mm, 5um)

Injection Volumn : 20ul sample loop

2. Chromatogram

