

Ginsenosides Rg1 Rb1 Analysis

1. Analytical Condition

YL9100 Series HPLC

Mobile Phase : ACN:Water=Gradient

Flow Rate : 1.0ml/min

Column Oven : 40°C

Detector : UV/Vis 203nm

Column : C18 (4.6*150mm, 5um)

Injection Volume : 20ul sample loop

Gradient Program	ACN	Water
Initial	15	85
15.00	15	85
45.00	40	60
50.00	50	50
55.00	100	0
60.00	100	0
62.00	15	85

2. Chromatogram

