

# Tert-Buthyl Pitavastatin (Cas No. 586966-54-3) Analysis

## 1. Analytical Condition

### YL9100 HPLC

Mobile Phase : ACN: Water = 60:40(in one bottle)

Flow Rate : 1ml/min

Column Oven : 40°C

Detector : UV/Vis 245nm

Column : MG C18 (4.6\*150mm, 5um)

Injection Volume : 10ul

## 2. Chromatogram

