

BTX (Benzene, Toluene, Xylene)

1. Analytical Condition

YL Gas Analyzer

Column: GS-Gaspro (15 m * 0.32 mm)

Oven: 50 °C (3min) → 20 °C/min → 250 °C (3min)

Carrier gas: He, 3mL/min (Split ratio 20 : 1)

Injector: Capillary, 230 °C

Detector: FID, 300 °C

Injection volume: 0.2 uL

2. Chromatogram

