


5-Aminotetrazole Analysis



1. Analytical Condition

YL9100 HPLC

Mobile Phase : A : Water 100% (add 0.1% phosphoric acid)

B : ACN 100% (add 0.1% phosphoric acid)

Flow Rate : 1 ml/min

Column Oven : 40°C

Detector : UV/Vis 240 nm

Column : Capcellpak MG C18 (4.6*250mm, 5um)

Injection Volume : 10ul

< Gradient Program >

Time (min)	A(%)	B(%)
0	100	0
5	50	50
10	50	50
10.1	100	0

2. Chromatogram

