

# PCB-138 Analysis

## 1. Analytical Condition

### YL 6500 Series

Oven : 150°C(1min)->10°C/min->300°C(1min)

Column : RTX-1 (30m\*0.32mm\*0.25um)

Carrier gas : N<sub>2</sub>, 3ml/min (Split ratio 5:1)

Injector : Capillary 330°C

Detector : ECD 330°C (Current=2.2, ATTN=1, Mkup 20ml/min)

Injection Volume : 2ul

## 2. Chromatogram

