

# Chitooligosaccharide Analysis

## 1. Analytical Condition

### YL 9100 Series GPC

Sample : Chitooligosaccharide(< M.W 2000 Chitosan) 1%

Mobile Phase : 0.02M NaNO<sub>3</sub>

Flow rate : 0.8 ml/min

Column oven : 35°C

Detector : RI(35°C)

Column : Ultrahydrogel 1000 + 120 (7.8 × 300 mm)

## 2. Chromatogram

