

## Propane, n-Butane, iso-Butane

### 1. Analytical Condition

#### YL Gas Analyzer

Carrier : He

Injector : 150°C, SR 1/10

Valve : 120 °C, 250 µL

Column : GS-Q (30.00 m \* 0.32 mm)

Oven : 100 °C(0.1 min) → 20 °C/min → 180 °C(1 min)

Detector : TCD, 150 °C, SEN 9

### 2. Chromatogram

