

# Acetaldehyde, Alcohols Analysis In Blood

## 1. Analytical Condition

### YL 6500 Series

Oven : 40°C(5min)->20°C/min->100°C(1min)

Column : IC Cap-Wax(30m\*0.53mm\*1.0um)

Carrier gas : N<sub>2</sub>, 6ml/min (Split ratio 7:1)

Injector : Capillary 230°C

Detector : FID 230°C

Headspace : 80 °C, 10min heating, Gas 1ml injection

## 2. Chromatogram

