


Chloroacetic Acid Analysis



1. Analytical Condition

YL9100 HPLC

Mobile Phase : Water:ACN=9.1(Add 0.15% phosphoric acid)

Flow Rate : 1ml/min

Column Oven : 45°C

Detector : UV/Vis 210nm

Column : Sunfire C18(4.6*150mm, 3.5um)

Injection Volume : 20ul

2. Chromatogram

