

General Solvents Analysis

1. Analytical Condition

YL 6500 Series

Oven : 150°C(1min)->5°C/min->170°C(1min)

Column : HP-INNOWAX(30m*0.53mm*1.0um)

Carrier gas : He, 7ml/min (Split ratio 5:1)

Injector : Capillary 250°C

Detector : FID 250°C

Sample preconditioning : 200 °C, 10min heating, Gas 1ml injection

2. Chromatogram

