

BTX(benzene, toluene, xylene) Analysis

1. Analytical Codntion

YL 6500 Series GC

Mobile phase : N₂ / 1 mL (SR 100:1)

Injector : 250 °C

Oven : 40°C (10 min) -> 20 °C(min) -> 200 °C(1min)

Detector : FID 250 °C

Column : RTX-2330 (60 m × 0.25 mm × 0.25 um)

Solvent : Hexane

2. Chromatogram

