



Antioxidants in Sesame Analysis

1. Analytical Condition

YL9100 HPLC System

Mobile Phase : Methanol : Water = 80 : 20 (1L, Premix)

Flow rate : 0.7 mL/min

Column : Shiseido C18 MG (4.6 * 250 mm, 5 μ m)

Detector : UV/Vis, 285 nm

Column Temp. : 35 $^{\circ}$ C

Injection Volume: 20 μ L

2. Chromatogram

